

MISSOURI DEPARTMENT OF HEALTH STATE PUBLIC HEALTH LABORATORY

RECEIVED DHSS Brouth Alcohol Program
By Carol Day at 12:22 pm, Mar 09, 2010

DATAMASTER MAINTENANCE REPORT

is repaired. Send copy to Department of Health; retain o	original in department file.				
DATAMASTER SN 200208		DATE OF INSPECTION 03-05-2010			
LOCATION OF INSTRUMENT (STREET AND CITY) 37 WEIS AVE, ELLISVILLE, MO. 63011	***************************************		NSPECTION		
CHECKLIST: Place a check (<) to the left of each item in observed values where determined.) Unchecked items	n if found to be satisfactory s must be corrected before u	or if operating sing instrume	within e	stablished limi	ts. (Write
DIAGNOSTIC CHECK (PRINTOUT ATTACHED)					
☑ COMPUTER	☑ DETECTOR	☑ DETECTOR			
	☑ FILTERS	☑ FILTERS			
☑ HEATERS SAMPLE CHAMBER 49 2C	☑ QUARTZ STAI	NDARD			
☑ FLOW DETECTOR	☑ CALIBRATION				
☑ PUMP HIGH SPEED	2 PRINTER				
☑ INDICATOR LIGHTS					
☑ TIME AND DATE					
SIMULATOR TEMPERATURE (34 °C ± 0.2°C) 34.0	00*C				
RECIRCULATION PUMP) Ø 0.100% STANDARD - MUST READ BETWEEN 0. 0.040% STANDARD - MUST READ BETWEEN 0. (ONLY ONE STANDARD IS TO BE USED PER MAIL	.038% and 0.042% INCLUSIV				
TEST 1 .096 TEST 2	.098	TEST 3	1 1 1 1 1 1 1 1 1 1	.099	
PERFORM R.F.I. TEST (PRINTOUT ATTACHED)					
M NUMBER OF REFUSALS, SINCE LAST MAINTENAN FOLLOWS: (DO NOT INCLUDE SIMULATOR TES	NCE REPORT, AND NUMBER	R OF BREATH	TESTS	IN EACH RAN	GE AS
	NCE REPORT, AND NUMBER	(.1519)	TESTS 2	IN EACH RANG	GE AS
FOLLOWS: (DO NOT INCLUDE SIMULATOR TEST APPLICATION OF THE CONTROL	(.1014) 0 (fication that was made to restauration) (ATION: .100	(.1519)	2	(Over .19)	0
FOLLOWS: (DO NOT INCLUDE SIMULATOR TEST REFUSALS 0 (004) 0 (.0509) List any new parts and describe any alteration or modified within established limits (use other side if necessary) SUPPLIER; REPCO MARKETING CONCENTR	(.1014) 0 (fication that was made to restauration) (ATION: .100	(.1519)	2	(Over .19)	0

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.

LOT NUMBER: 09002

EXPIRATION DATE: August 31, 2011 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number <u>09002</u> of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain <u>.1206</u> gms/dl +/-.003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of <u>.100</u> +/-3% gms/210L Breath when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is <u>September 1, 2009.</u>

The expiration date for this lot number is <u>August 31, 2011</u> at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.

Cecil B. Garner, President RepCo Marking, Inc.

Form RM 02

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster

Evidence Ticket

ELLISVILLE POLICE DEPARTMENT STATE OF MISSOURI

BAC DATAMASTER SERIAL NUMBER 200208 00/05/10 0

PERMIT NUMBER: SESESE OFFICER I.D.: 189 TESTING OFFICER: DIEHL/JEFF

--- SUPERVISOR MODE

WINCELLENGOUS DATA:

EXPIRETION BETE: 00/00/10

BLUZK TENT	EXTERNET WHENDERD	BCDZX TOOT	MXTMRZEL WTRABERD	BCBZX TMOT	EXTERNAL STANDARD	INTERNAL STANDARD	BLANK TEST
න න න	000	200	, 10 10	, 2000 ,	, 50 50 50		2000 2000
91.14 0	四 1.40	空 4 4	©::41	四二:40	68:18	00 10 10 10 10 10 10 10 10 10 10 10 10 1	91:39

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster

Evidence Ticket

ELLISVILLE POLICE DEPARTMENT STATE OF MISSOURI

BAC INTAMASTER SERIAL NUMBER 200208 00/05/10 80 H H 10 G

BIAGNOSTIC CHECK ---

COMPUTER: 989

PROGRAM <04-07-0009>: 9897

THRITHRO

SHAPLE CHAMBER! 49 10 11

FLOW DETECTOR: OKAY

PUMB

HIGH SPEED:

2620

THITTOTOR: 9KBY

FILTERS:

OKAY

z II

EIN. AVG. =

. 20070

QUARTZ STANDARD: 2625

CHLIBRATION: 吴号

Pqrstuvwxyz<1>+ HIJKLMMOPQRSTUVWXYZ[\]^_^abcdefghijklmno !"#\$%%;()*+,-,/0100456789:;(=>?@ABCDEFG PRINTER TEST

OPERATOR SIGNATURE

REORDER ALL SUPPLIES FROM N.P.A.S.

80. But 185

CARD STK # OPERATOR SIGNATURE

REORDER ALL SUPPLIES FROM N.P.A.S.

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster

Evidence Ticket

STATE OF MISSOURI

BAC DATAMASTER SERIAL NUMBER 200208 03/05/10

ARREST TIME: 00:01 SUBJECT NAME:

RFITEST

DOB: 11/11/11 SHOW IN

STATE/D.L.: MO/11111 ARRESTING OFFICER: RFITEST

OFFICER I.D.: RFI TESTING OFFICER:

DIEHL/JEFF

OFFICER I.D.: 189 PERMIT NUMBER: 820252

EXPIRATION DATE: 89/82/18 MISCELLANEOUS DATA:

RFI TEST

--- BREATH ANALYSIS ---

RADIO INTERFERENCE

OPERATOR SIGNATURE

REORDER ALL SUPPLIES FROM N.P.A.S.

CARD STK #

State of Missouri DEPARTMENT OF HEALTH



PERMIT TYPE II



JEFF DIEHL				
is hereby authorized to instruct a calibrate, perform field repairs, and	and supervise operators, train instructors, inspect, operate the following breath analyzer(s):			
DATAMA	STER			
for the determination of the alcoholic air. Issued under the provisions of se	c content of blood from a sample of expired (alveolar) ections 577.020 through 577.041, RSMo 1986.			
09/02/08	Time C. Oo Cont			
820252	Director of State Public Health Laboratory			
09/02/2010	from the second of			
Expires				
MO 580-0771 (7-88)	- Director, Department of Health Lab. 4 (R7-88)			